

**Title:**

**Application of an electronic tongue to detect gliadins in gluten-free foods**

**Authors & affiliations:**

A.M. Peres<sup>1,2</sup>, L.G. Dias<sup>2</sup>, A.C.A. Veloso<sup>3,4</sup>, S.G. Meirinho<sup>2</sup>, J. Sá Morais<sup>2</sup>, A.A.S.C. Machado<sup>5</sup>

<sup>1</sup>LSRE - Escola Superior Agrária de Bragança, Portugal

<sup>2</sup>CIMO - Escola Superior Agrária de Bragança, Portugal

<sup>3</sup>DEQB - Instituto Superior de Engenharia de Coimbra, Portugal

<sup>4</sup>IBB - Center of Biological Engineering, University of Minho, Portugal

<sup>5</sup>LAQUIPAI - Departamento de Química, Faculdade de Ciências da Universidade do Porto, Portugal

**Abstract:**


Celiac disease is an autoimmune-mediated disorder triggered in genetically susceptible individuals by the ingestion of some gluten proteins, namely gliadins. To prevent inadvertent gluten consumption, the Commission Regulation (EC) N°41/2009 will implement labelling foods as “gluten-free” or “low-gluten content”[1].

The feasibility of an all-solid-state potentiometric electronic tongue to detect the contamination of “gluten-free” foodstuffs with gliadins was evaluated. The device was constituted by 36 cross-sensitivity lipo/polymeric membranes and its performance was assessed using food ethanolic extracts (Figure 1).


A semi-quantitative linear discriminant (LDA) model was established based on the signal patterns of 11 polymeric membranes. The model was able to distinguish between “Gluten-free” extracts (<10 ppm gliadins), “Low-Gluten content” extracts (20-50 ppm gliadins) and “Gluten-containing” extracts (>50 ppm gliadins), with sensibilities of 100% and 77% for the original grouped cases and “leaving one-out” cross-validation, respectively (Figure 2). The device sensitivity was towards 1-2 ppm.

The device was also applied to discriminate between “gluten-free” and “gluten-containing” foodstuffs (flours, baby milked flours, breads, cookies and breakfast cereals), being the label information checked by HPLC-DAD analysis. In this case, the LDA was based on the signals of only 4 polymeric membranes. The model allowed to classify correctly 89% and 84%, for the original grouped cases and “leaving one-out” cross-validation procedures, respectively, which is very satisfactory taking in account the higher complexity due to the food matrix effect. Furthermore, only one of the “Gluten-containing” foodstuffs (a bread sample) was misclassified by the model as “Gluten-free”. This misclassification could be tentatively due to an additional difficulty in the gliadins extraction from bread, which is a more processed food matrix.

Finally, the work carried out showed that the E-tongue device could be used in practice as a fast and economic preliminary tool to evaluate, in a real time basis, the possible gluten contaminations of “Gluten-free” foodstuffs.



*Fig. 1: Experimental set-up: electronic tongue*



*Fig. 2: Linear discriminant analysis classification based on the signal profiles recorded by the electronic tongue.*

## References

- [1] Commission Regulation (EC) N°41/2009 of 20 January 2009 concerning the composition and labelling of foodstuffs suitable for people intolerant to gluten, Official Journal of the European Union of 21.1.2009, L 16/3-L 16/5.